



HAL
open science

A glucose electrode using high-stability glucose-oxidase collagen membranes

Daniel R. Thevenot, Robert Sternberg, Pierre Coulet

► **To cite this version:**

Daniel R. Thevenot, Robert Sternberg, Pierre Coulet. A glucose electrode using high-stability glucose-oxidase collagen membranes. *Diabetes Care*, 1982, 5 (3), pp.203 - 206. 10.2337/diacare.5.3.203 . hal-01083014

HAL Id: hal-01083014

<https://enpc.hal.science/hal-01083014>

Submitted on 14 Nov 2014

HAL is a multi-disciplinary open access archive for the deposit and dissemination of scientific research documents, whether they are published or not. The documents may come from teaching and research institutions in France or abroad, or from public or private research centers.

L'archive ouverte pluridisciplinaire **HAL**, est destinée au dépôt et à la diffusion de documents scientifiques de niveau recherche, publiés ou non, émanant des établissements d'enseignement et de recherche français ou étrangers, des laboratoires publics ou privés.



Distributed under a Creative Commons Attribution 4.0 International License

A Glucose Electrode Using High-Stability Glucose-Oxidase Collagen Membranes

DANIEL R. THÉVENOT, ROBERT STERNBERG, AND PIERRE COULET

A very sensitive glucose electrode was developed using a glucose-oxidase membrane and an anodically polarized platinum disk. Calibration curves were linear over 4.5 concentration decades. It was adapted for human blood samples. The lifetime of glucose-oxidase collagen membranes was greater than 3 yr at 4°C and was ~6 mo at 20–30°C. DIABETES CARE 5: 203–206, MAY–JUNE 1982.

Among the various techniques available for designing a glucose sensor, the glucose electrode, which combines a glucose-oxidase membrane and an electrochemical detector, is certainly the most developed. Since the first report by Clark et al.¹ and Updike et al.,² different detectors have been tested and different methods of immobilization have been studied. These include entrapment of enzyme solution by semipermeable membrane or by gel, and coreticulation or covalent coupling.^{3,4} However, one of the main problems encountered with such enzyme electrodes has been the instability of enzymatic membranes. We have developed a sensor^{5–7} using glucose oxidase covalently bound to collagen by the acyl-azide procedure,^{8–10} which has excellent stability.¹¹ This article summarizes the main analytical characteristics of such a sensor, especially with blood samples, and presents the results of stability studies in storage and operating conditions.

MATERIAL AND METHODS

Glucose electrode. Glucose oxidase was coupled by the previously described acyl-azide procedure applied to insoluble films of highly polymerized reconstituted collagen^{8–10} prepared by the Centre Technique du Cuir (Lyon, France). Carboxyl groups were first esterified by immersion in a methanol-hydrochloric acid solution and then treated by hydrazine before reacting with nitrous acid. The acyl-azide groups of such activated collagen reacted spontaneously with a glucose-oxidase buffered solution.

As pointed out previously, we preferred to use the glucose-oxidase membrane in association with anodic detection of hydrogen peroxide so that the glucose electrode is less dependent on oxygen concentration and so that its output current

increases with the glucose level.⁴ Since such detection is not very selective to enzymatically generated hydrogen peroxide, we included in our device a compensating electrode with a nonenzymatic collagen membrane. Except for data processing, material and methods were identical to those previously described.^{5–7}

Electrode responses. When a glucose-containing sample is added to a solution of 0.2 M acetate buffer \pm 0.1 M KCl (pH 5.7) into which both electrodes are dipped, three different current-versus-time curves may be recorded:

(1) I_2 is the output current of the nonenzymatic, compensating electrode, E_2 . I_2 is the background response and is usually very low except if the sample contains electrochemical reducing species such as ascorbate, urate, or sulfites.

(2) $(I_1 - k \cdot I_2)$, where k is close to 1, and corresponds to the detection of enzymatically generated hydrogen peroxide; it reaches a steady-state value after 2–3 min; this is the steady-state response of the sensor.

(3) $d(I_1 - k \cdot I_2)/dt$ is maximum after 30–50 s; the height of this peak is the dynamic response of the sensor.

Data processing. In order to avoid time-consuming analyses of recordings, we added to the previous equipment a Hewlett-Packard 97S programmable table calculator. Its binary coded decimal inputs and outputs were interfaced to the potentiostat and to an electronic buret through a Solea-Tacusel Ionomate 80 digital millivoltmeter (Figure 1). We developed a program that performs two main functions: (1) determination of steady-state response by detecting stable output currents before and after the sample addition, and (2) control of glucose standard additions by the electronic buret. We found that such a device gave printed results and a precision that was equal to graphical analysis of analogue recordings, providing we fixed low enough stability criteria.




FIG. 3. Stability of steady-state responses during operation with seven different glucose-oxidase collagen membranes. All membranes were tested at 30°C and stored at 4°C, except (●), which was stored at room temperature.

ure 3. Thus we modified our storage procedure so that each freshly coupled membrane was stored at 4°C until it was first mounted on an electrode body; then it was kept at room temperature for several months, pressed against a polarized platinum disk, and used for more than 1000 assays at 30°C, accumulating 1000 h of operation. The evolution of daily or weekly calibrations for three membranes that had undergone various periods of storage at 4°C is presented in Figure 4. Two observations can be made from these results. First, the responses obtained with the same membrane may vary significantly over weeks or months. This does not seem to be re-




FIG. 4. Stability of steady-state responses when after (○) 1, (▲) 229, and (■) 755 days at 4°C. The glucose-oxidase collagen membranes were stored at room temperature and operated at 30°C (■). Data were obtained from ref. 13.

lated to poor regulation of the cell temperature, but more probably results from slight variations of geometrical parameters of the electrodes such as platinum disk-membrane distance. Indeed, it was found that an increase of the pressure of this enzymatic membrane against the platinum disk yields an important increase of steady-state responses and an even more important increase in dynamic responses.¹² Second, the responses do not decrease significantly after more than 2 yr at 4°C and 100 or even 200 days at 20–30°C; for such membranes, the surface activity of which ranged between 40 and 60 mU/cm², we did not find any correlation between the loss of activity of the enzymatic membranes and their operation time in contact with glucose-containing solutions, such as blood, food, or river water samples.¹²

DISCUSSION AND CONCLUSION

It is generally claimed that enzyme immobilization, and especially its covalent coupling to an insoluble carrier, stabilizes the enzyme activity. Klibanov¹⁴ recently reviewed the different factors to which the apparent stability is related.

When the enzymatic load is large enough and thus its intrinsic activity high compared with mass transfer kinetics, the rate-limiting factor is substrate diffusion and the intrinsic activity cannot be observed. Then, the overall reaction rate is minimally affected or not affected at all by a partial deactivation of the immobilized enzyme.¹⁵ Several illustrations of such behavior have been given by Racine¹⁶ or Comtat and Mahenc^{17–19} using highly concentrated enzyme solutions entrapped by dialysis film, and by Scheller²⁰ or Gondo²¹ using highly concentrated enzyme gels.

By contrast, the glucose-oxidase collagen membranes used in this glucose sensor seem to be kinetically limited by the enzymatic reaction.^{6,22} A partial deactivation of the coupled glucose oxidase would directly affect the overall surface activity of the membrane and the electrode response. The results presented in this report together with those obtained with collagen-bound aspartate aminotransferase¹¹ show that even if diffusional effects can partially mask the intrinsic behavior of bound enzymes, the proteinaceous environment provided by collagen seems very favorable for storage and operational stability. Work is in progress to better define some geometric parameters of the glucose electrode, such as the thickness of the liquid film situated between the membrane and the platinum disk, used as a working electrode for hydrogen peroxide detection. Nevertheless, this glucose electrode can be used with daily calibration for several months at a time without replacing its high-stability glucose-oxidase collagen membrane.

From the Laboratoire de Bioélectrochimie et d'Analyse du Milieu, UER de Sciences, Université Paris Val de Marne, Avenue du Général de Gaulle F.94010 Creteil Cédex (DRT, RS), and the Laboratoire de Biologie et Technologie des Membranes du CNRS, Université Claude Bernard (Lyon I), 43, boulevard du 11 novembre 1918 F.69621 Villeurbanne (PC).

Address reprint requests to Daniel R. Thévenot, Laboratoire de Bioélectrochimie et d'Analyse du Milieu, UER de Sciences, Uni-

versité Paris Val de Marne, Avenue du Général de Gaulle F.94010 Creteil Cédex, France.

REFERENCES

- ¹ Clark, L. C., Jr., and Lyons, C.: Electrode systems for continuous monitoring in cardiovascular surgery. *Ann. N.Y. Acad. Sci.* 102: 29-49, 1962.
- ² Updike, S. J., and Hicks, G. P.: The enzyme electrode. *Nature* 214: 986-88, 1967.
- ³ Guilbault, G. G.: Enzyme electrode probes. In *Immobilized Enzymes, Antigen, Antibodies and Peptides*. Weetall, H. H., Ed. New York, Marcel Dekker, 1975, pp. 293-417.
- ⁴ Thévenot, D. R.: Problems in adapting a glucose-oxidase-electrochemical sensor into an implantable glucose sensing device. *Diabetes Care* 5: 184-89, 1982.
- ⁵ Thévenot, D. R., Coulet, P. R., Sternberg, R., and Gautheron, D. C.: A highly sensitive glucose electrode using glucose oxidase collagen film. *Bioelectrochem. Bioenergetics* 5: 548-53, 1978.
- ⁶ Thévenot, D. R., Sternberg, R., Coulet, P. R., Laurent, J., and Gautheron, D. C.: Enzyme collagen membrane for electrochemical determination of glucose. *Anal. Chem.* 51: 96-100, 1979.
- ⁷ Sternberg, R., Apoteker, A., and Thévenot, D. R.: Trace glucose electrode for clinical, food and environment determinations. *Anal. Chem. Symposia Ser. vol 2*. Amsterdam, Elsevier, pp. 461-73.
- ⁸ Coulet, P. R., Julliard, J. H., and Gautheron, D. C.: French patent 73-23-283, publ. no. 2, 235, 133, 1973.
- ⁹ Coulet, P. R., Julliard, J. H., and Gautheron, D. C.: A mild method of general use for covalent coupling of enzymes to chemically activated collagen films. *Biotechnol. Bioeng.* 16: 1055-68, 1974.
- ¹⁰ Coulet, P. R., and Gautheron, D. C.: Enzyme anchoring on chemically activated collagen membranes. In *Analysis and Control of Immobilized Enzyme Systems*. Thomas, D., and Kernevez, J. P., Eds. Amsterdam, North Holland, 1976, pp. 165-77.
- ¹¹ Coulet, P. R., and Gautheron, D. C.: Resistance to deactivation of enzyme-collagen membranes. *Biochimie* 62: 543-47, 1980.
- ¹² Sternberg, R.: Enzyme electrodes with enzymes coupled to collagen membrane: determination of glucose and urea. Créteil, Université Paris-Val de Marne, 1979. Thesis.
- ¹³ De Laforcade, V.: Study of an instrument. Application to glucose determination with enzyme electrode. Créteil Université Paris-Val de Marne, 1980. Diploma.
- ¹⁴ Klibanov, A. M.: Enzyme stabilization by immobilization. *Anal. Biochem.* 93: 1-25, 1979.
- ¹⁵ Thévenot, D., and Sternberg, R.: Rate limiting phenomena in enzyme electrodes. Communication at Journées d'Electrochimie, Strasbourg, 1979.
- ¹⁶ Racine, P., and Mindt, W.: On the role of substrate diffusion in enzyme electrodes. *Experientia (Suppl.)* 18: 525-34, 1971.
- ¹⁷ Durliat, H., Comtat, M., Mahenc, J., and Baudras, A.: Amperometric determination of lactate using enzyme electrode. *J. Electroanal. Chem.* 66: 73-76, 1975.
- ¹⁸ Durliat, H., Comtat, M., Mahenc, J., and Baudras, A.: Study of the optimal operating conditions of an enzyme electrode specific for lactate. Adaptation for blood titration. *Anal. Chim. Acta* 85: 31-40, 1976.
- ¹⁹ Mahenc, J., and Aussaresses, H.: Enzyme electrode specific for glucose using amperometric detection of hexacyanoferrate II. *C. R. Acad. Sci. Paris, Ser. C* 289: 357-59, 1979.
- ²⁰ Scheller, F., and Pfeiffer, D.: Enzyme electrodes. *Zeitschrift Chemie* 2: 50-57, 1978.
- ²¹ Gondo, S., Morishita, M., and Osaki, T.: Improvement of glucose sensor performance with immobilized glucose oxidase-glucose isomerase system. *Biotechnol. Bioeng.* 22: 12, 1980.
- ²² Coulet, P. R., Sternberg, R., and Thévenot, D. R.: Electrochemical study of reactions at interfaces of glucose oxidase collagen membranes. *Biochim. Biophys. Acta* 612: 317-27, 1980.